

Scope of Laboratory Accreditation

Laboratory Name : Center of Measurement and Standard Accreditation, Faculty of Science,
Prince of Songkla University

Address : 15 Karnjanavanit Road, Tambon Kho Hong, Amphoe Hat Yai,
Changwat Songkhla 90112

Accreditation Number : Testing - 0033

Laboratory Status : Permanent Site Temporary Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
1	Water	- Manganese 0.01 mg/L to 1.0 mg/L	In - house method : T-1 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 3120 B, 3030 E
2	Effluent	- Copper 0.05 mg/L to 1.0 mg/L - Iron 0.05 mg/L to 1.0 mg/L - Zinc 0.05 mg/L to 1.0 mg/L	In - house method : T-2 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 3120 B, 3030 E

Issue Date : 26th October 2023

Signature : 
(Mrs. Chantarat Vorasapavit)

Scientist, Senior Professional Level

Acting Director of Bureau of Laboratory Accreditation

Initial Issue Date 26th May 2009

Issue Number 9